



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

June 10, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	22JUSFSY059	JUN.09 2023

The product kit, batch 22JUSFSY059, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : June 10, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.006	0.006	0.006	0.005	0.006	
0.78	0.047	0.048	0.048	0.047	0.048	1.2
1.56	0.090	0.089	0.090	0.087	0.089	1.6
3.13	0.169	0.167	0.165	0.162	0.166	1.8
6.25	0.316	0.314	0.309	0.304	0.311	1.7
12.5	0.606	0.584	0.595	0.589	0.594	1.6
25	1.099	1.079	1.109	1.090	1.094	1.2
50	1.985	1.999	2.018	1.962	1.991	1.2

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.9	6.2-10.1
B	9.0	8.5-13.8
C	5.5	4.4-7.8