



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

May 19, 2022

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

### Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	22MYSFSY058	MAY.18 2023

The product kit, batch 22MYSFSY058, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : May 19, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.008	0.007	0.008	0.007	0.008	
0.78	0.050	0.049	0.053	0.048	0.050	4.3
1.56	0.092	0.086	0.093	0.088	0.090	3.7
3.13	0.169	0.172	0.169	0.166	0.169	1.4
6.25	0.320	0.319	0.322	0.316	0.319	0.8
12.5	0.601	0.605	0.610	0.606	0.606	0.6
25	1.056	1.152	1.136	1.148	1.123	4.0
50	2.014	2.043	2.081	2.018	2.039	1.5

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.8	6.2-10.1
B	10.6	8.5-13.8
C	6.0	4.4-7.8