



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

March 24, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	22MASFSY057	MAR.23 2023

The product kit, batch 22MASFSY057, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : March 24, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.016	0.015	0.016	0.014	0.015	
0.78	0.061	0.057	0.057	0.056	0.058	3.8
1.56	0.096	0.096	0.095	0.095	0.096	0.6
3.13	0.171	0.172	0.168	0.166	0.169	1.6
6.25	0.322	0.316	0.306	0.301	0.311	3.1
12.5	0.603	0.561	0.556	0.557	0.569	4.0
25	1.114	1.051	1.052	1.062	1.070	2.8
50	2.022	1.986	1.996	1.974	1.995	1.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.1	6.3-11.7
B	12.2	8.8-16.3
C	6.8	4.3-8.1