

**Morinaga Institute of Biological Science, Inc.**

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

January 14, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	22JASFSY055	JAN.13 2023

The product kit, batch 22JASFSY055, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : January 14, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.032	0.032	0.031	0.029	0.031	
0.78	0.077	0.076	0.075	0.072	0.075	2.9
1.56	0.116	0.111	0.109	0.108	0.111	3.2
3.13	0.206	0.197	0.188	0.184	0.194	5.1
6.25	0.367	0.349	0.335	0.340	0.348	4.0
12.5	0.648	0.631	0.620	0.607	0.627	2.8
25	1.221	1.165	1.145	1.152	1.171	2.9
50	2.185	2.190	2.103	2.120	2.150	2.1

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.8	6.3-11.7
B	14.1	8.8-16.3
C	7.0	4.3-8.1