



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

May 13, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	21MYSFSY051	MAY.12 2022

The product kit, batch 21MYSFSY051, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : May 13, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.008	0.007	0.009	0.008	0.008	
0.78	0.044	0.043	0.041	0.042	0.043	3.0
1.56	0.075	0.074	0.072	0.072	0.073	2.0
3.13	0.144	0.141	0.138	0.139	0.141	1.9
6.25	0.273	0.273	0.265	0.265	0.269	1.7
12.5	0.530	0.520	0.515	0.496	0.515	2.8
25	0.990	0.973	0.966	0.946	0.969	1.9
50	1.794	1.767	1.770	1.721	1.763	1.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.2	6.3-11.7
B	11.9	8.8-16.3
C	6.9	4.3-8.1