



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

November 19, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	20NOSFSY046	NOV.18 2021

The product kit, batch 20NOSFSY046, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : November 19, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.044	0.043	0.044	0.044	0.044	
0.78	0.099	0.098	0.096	0.099	0.098	1.4
1.56	0.148	0.148	0.146	0.146	0.147	0.8
3.12	0.246	0.242	0.247	0.246	0.245	0.9
6.25	0.440	0.438	0.438	0.438	0.439	0.2
12.5	0.803	0.796	0.784	0.783	0.792	1.2
25	1.420	1.417	1.405	1.395	1.409	0.8
50	2.321	2.378	2.305	2.394	2.350	1.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.9	6.3-11.7
B	12.4	8.8-16.3
C	6.7	4.3-8.1