

September 17, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	20SESFSY045	SEP.16 2021

The product kit, batch 20SESFSY045, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : September 17, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.008	0.008	0.009	0.009	0.009	
0.78	0.052	0.051	0.051	0.052	0.052	1.1
1.56	0.094	0.095	0.094	0.091	0.094	1.9
3.12	0.178	0.180	0.178	0.173	0.177	1.7
6.25	0.345	0.342	0.346	0.334	0.342	1.6
12.5	0.666	0.650	0.658	0.632	0.652	2.2
25	1.229	1.236	1.187	1.184	1.209	2.3
50	2.163	2.185	2.062	2.121	2.133	2.5

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.9	6.3-11.7
B	12.1	8.8-16.3
C	6.7	4.3-8.1