

**Morinaga Institute of Biological Science, Inc.**

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

March 13, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	20MASFSY041	MAR.12 2021

The product kit, batch 20MASFSY041, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : March 13, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.007	0.009	0.008	0.008	0.008	
0.78	0.048	0.052	0.051	0.053	0.051	4.2
1.56	0.086	0.087	0.087	0.091	0.088	2.5
3.12	0.161	0.161	0.156	0.164	0.161	2.1
6.25	0.313	0.309	0.306	0.315	0.311	1.3
12.5	0.580	0.596	0.561	0.588	0.581	2.6
25	1.108	1.118	1.113	1.110	1.112	0.4
50	1.924	1.923	1.866	1.945	1.915	1.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.1	6.3-11.7
B	11.4	8.8-16.3
C	6.2	4.3-8.1