

**Morinaga Institute of Biological Science, Inc.**

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

November 18, 2019

## Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

**Product:**

Product name	Batch number	Expiration date
Soya ELISA Kit II	19NOSFSY039	NOV.14 2020

The product kit, batch 19NOSFSY039, passed the quality control.

**(1) Absorbance of the standard curve**

(Assay date : November 18, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.045	0.042	0.042	0.041	0.043	4.1
0.78	0.085	0.085	0.081	0.081	0.083	2.8
1.56	0.121	0.117	0.116	0.115	0.117	2.2
3.12	0.192	0.185	0.181	0.182	0.185	2.7
6.25	0.342	0.323	0.321	0.328	0.329	2.9
12.5	0.605	0.563	0.578	0.572	0.580	3.1
25	1.123	1.076	1.086	1.070	1.089	2.2
50	2.046	2.013	2.001	2.008	2.017	1.0

**(2) Results of the Test materials**

Positive Control	Protein content(μg/g)	Standard value
A	11.3	9.2-17.0
B	11.7	8.9-16.5
C	6.4	4.8-8.8