

October 16, 2019

## Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



**Product:**

Product name	Batch number	Expiration date
Soya ELISA Kit II	19OCSFSY038	OCT.15 2020

The product kit, batch 19OCSFSY038, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : October 16, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.008	0.011	0.009	0.010	13.6
0.78	0.053	0.056	0.056	0.054	0.055	2.7
1.56	0.102	0.102	0.103	0.100	0.102	1.2
3.12	0.187	0.186	0.185	0.181	0.185	1.4
6.25	0.349	0.364	0.362	0.343	0.355	2.9
12.5	0.677	0.682	0.678	0.664	0.675	1.2
25	1.241	1.251	1.250	1.229	1.243	0.8
50	2.187	2.156	2.185	2.146	2.169	1.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.8	6.3-11.7
B	10.6	8.8-16.3
C	6.5	4.3-8.1