

August 30, 2019

## Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



**Product:**

Product name	Batch number	Expiration date
Soya ELISA Kit II	19AUSFSY037	AUG.28 2020

The product kit, batch 19AUSFSY037, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : August 30, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.011	0.011	0.012	0.012	5.0
0.78	0.054	0.055	0.053	0.053	0.054	1.8
1.56	0.094	0.096	0.092	0.092	0.094	2.0
3.12	0.169	0.171	0.171	0.165	0.169	1.7
6.25	0.320	0.308	0.320	0.313	0.315	1.9
12.5	0.609	0.600	0.574	0.578	0.590	2.9
25	1.119	1.115	1.092	1.071	1.099	2.0
50	2.011	2.003	2.026	1.959	2.000	1.4

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.3	6.3-11.7
B	11.8	8.8-16.3
C	7.2	4.3-8.1