

**Morinaga Institute of Biological Science, Inc.**

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

June 14, 2019

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	19JUSFSY035	JUN.12 2020

The product kit, batch 19JUSFSY035, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : June 14, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.062	0.059	0.063	0.061	0.061	2.8
0.78	0.105	0.102	0.107	0.107	0.105	2.2
1.56	0.133	0.134	0.135	0.155	0.139	7.6
3.12	0.199	0.196	0.204	0.199	0.200	1.7
6.25	0.322	0.325	0.325	0.331	0.326	1.2
12.5	0.562	0.576	0.555	0.560	0.563	1.6
25	1.036	1.022	1.024	1.047	1.032	1.1
50	1.919	1.927	1.906	1.925	1.919	0.5

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.5	9.2-17.0
B	12.0	8.9-16.5
C	7.0	4.8-8.8