

**Morinaga Institute of Biological Science, Inc.**

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

April 10, 2019

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Soya ELISA Kit II	19APSFSY033	APR.08 2020

The product kit, batch 19APSFSY033, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : April 10, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.011	0.012	0.014	0.010	0.012	14.5
0.78	0.055	0.055	0.053	0.052	0.054	2.8
1.56	0.092	0.096	0.095	0.098	0.095	2.6
3.12	0.169	0.174	0.174	0.176	0.173	1.7
6.25	0.325	0.336	0.338	0.347	0.337	2.7
12.5	0.600	0.615	0.618	0.629	0.616	1.9
25	1.117	1.148	1.113	1.142	1.130	1.6
50	1.832	1.959	1.957	1.993	1.935	3.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.4	6.3-11.7
B	10.5	8.8-16.3
C	5.5	4.3-8.1