



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

February 19, 2024

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Sesame ELISA Kit II	24FESFSE009	FEB.15 2025

The product kit, batch 24FESFSE009, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : February 19, 2024)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.002	0.002	0.002	0.002	0.002	
0.39	0.043	0.043	0.040	0.041	0.042	3.6
0.78	0.085	0.085	0.085	0.085	0.085	0.0
1.56	0.166	0.167	0.165	0.164	0.166	0.8
3.13	0.324	0.335	0.323	0.329	0.328	1.7
6.3	0.631	0.645	0.635	0.632	0.636	1.0
12.5	1.212	1.190	1.188	1.132	1.181	2.9
25	1.988	2.050	2.040	2.026	2.026	1.3

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.5	4.8-8.8
B	3.5	2.2-4.0
C	5.0	3.3-6.2