



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

September 7, 2022

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

### Product:

Product name	Batch number	Expiration date
Sesame ELISA Kit II	22SESFSE005	SEP.06 2023

The product kit, batch 22SESFSE005, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : September 7, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.003	0.004	0.004	0.004	0.004	
0.39	0.044	0.045	0.046	0.045	0.045	1.8
0.78	0.091	0.091	0.092	0.089	0.091	1.4
1.56	0.174	0.176	0.177	0.175	0.176	0.7
3.13	0.336	0.340	0.340	0.337	0.338	0.6
6.3	0.662	0.661	0.659	0.647	0.657	1.1
12.5	1.237	1.243	1.213	1.197	1.223	1.7
25	2.175	2.171	2.108	2.005	2.115	3.8

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.8	4.8-8.8
B	3.2	2.2-4.0
C	5.1	3.3-6.2