



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

June 14, 2022

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

### Product:

Product name	Batch number	Expiration date
Sesame ELISA Kit II	22JUSFSE004	JUN.13 2023

The product kit, batch 22JUSFSE004, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : June 14, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.003	0.003	0.004	0.003	0.003	
0.39	0.043	0.040	0.043	0.041	0.042	3.6
0.78	0.078	0.077	0.079	0.079	0.078	1.2
1.56	0.147	0.147	0.152	0.152	0.150	1.9
3.13	0.284	0.284	0.296	0.300	0.291	2.8
6.3	0.545	0.551	0.556	0.564	0.554	1.5
12.5	1.039	1.064	1.078	1.081	1.066	1.8
25	1.888	1.898	1.962	1.970	1.930	2.2

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	6.5	4.8-8.8
B	2.9	2.2-4.0
C	4.8	3.3-6.2