



## Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,  
YOKOHAMA, 2360003, JAPAN  
TEL:+81-45-791-7674  
E-mail: info\_miobs\_e@morinaga.co.jp

March 15, 2022

# Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

### Product:

Product name	Batch number	Expiration date
Sesame ELISA Kit II	22MASFSE003	MAR.14 2023

The product kit, batch 22MASFSE003, passed the quality control.

### (1) Absorbance of the standard curve

(Assay date : March 15, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.006	0.007	0.006	0.006	0.006	
0.39	0.050	0.049	0.048	0.046	0.048	3.5
0.78	0.095	0.093	0.089	0.087	0.091	4.0
1.56	0.183	0.180	0.173	0.168	0.176	3.9
3.13	0.354	0.338	0.335	0.329	0.339	3.1
6.3	0.666	0.642	0.624	0.626	0.640	3.0
12.5	1.173	1.204	1.176	1.168	1.180	1.4
25	2.062	2.180	2.053	2.099	2.099	2.8

### (2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.7	4.6-9.6
B	3.5	2.2-4.6
C	6.1	3.3-7.5