



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

July 28, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Sesame ELISA Kit II	21JLSFSE002	JUL.27 2022

The product kit, batch 21JLSFSE002, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : July 28, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.006	0.008	0.008	0.006	0.006	
0.39	0.046	0.045	0.047	0.044	0.046	2.8
0.78	0.083	0.081	0.087	0.082	0.083	3.2
1.56	0.157	0.152	0.172	0.159	0.157	5.3
3.13	0.308	0.304	0.331	0.308	0.308	3.9
6.3	0.587	0.576	0.627	0.591	0.587	3.7
12.5	1.094	1.112	1.177	1.117	1.094	3.2
25	2.029	2.042	2.180	2.056	2.029	3.4

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.0	4.6-9.6
B	3.4	2.2-4.6
C	5.0	3.3-7.5