



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

February 14, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Crustacean ELISA Kit II	23FESFCR008	FEB.12 2024

The product kit, batch 23FESFCR008, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : February 14, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.016	0.017	0.015	0.014	0.016	
0.78	0.048	0.051	0.049	0.047	0.049	3.5
1.56	0.087	0.095	0.089	0.083	0.089	5.6
3.13	0.159	0.167	0.155	0.145	0.157	5.8
6.25	0.291	0.283	0.295	0.273	0.286	3.4
12.5	0.564	0.557	0.572	0.574	0.567	1.4
25	1.099	1.112	1.111	1.125	1.112	1.0
50	2.090	2.195	2.154	2.138	2.144	2.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.6	5.7-10.6
B	10.0	6.9-12.8
C	9.7	6.8-12.7