



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

October 20, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Crustacean ELISA Kit II	20OCSFCR003	OCT. 18 2021

The product kit, batch 20OCSFCR003, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : October 20, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.032	0.030	0.030	0.029	0.030	
0.78	0.073	0.069	0.068	0.068	0.070	3.4
1.56	0.112	0.102	0.103	0.105	0.106	4.3
3.13	0.189	0.175	0.173	0.172	0.177	4.5
6.25	0.341	0.316	0.307	0.321	0.321	4.5
12.5	0.651	0.580	0.587	0.597	0.604	5.3
25	1.238	1.156	1.170	1.166	1.183	3.2
50	2.280	2.233	2.122	2.237	2.218	3.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.8	5.7-10.6
B	10.3	6.9-12.8
C	9.2	6.8-12.7