

June 06, 2019

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:



Product:

Product name	Batch number	Expiration date
Crustacean ELISA Kit II	19JUSFCR001	JUN. 05 2020

The product kit, batch 19JUSFCR001, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : June 06, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.019	0.017	0.016	0.016	0.017	8.3
0.78	0.047	0.046	0.045	0.044	0.046	2.8
1.56	0.078	0.075	0.073	0.073	0.075	3.2
3.12	0.148	0.138	0.139	0.140	0.141	3.2
6.25	0.262	0.256	0.266	0.261	0.261	1.6
12.5	0.517	0.505	0.515	0.497	0.509	1.8
25	0.960	0.950	0.988	0.939	0.959	2.2
50	1.848	1.823	1.799	1.770	1.810	1.8

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	8.1	5.7-10.6
B	10.4	6.9-12.8
C	9.4	6.8-12.7