



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL: +81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

December 5, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Monta

Product:

Product name	Batch number	Expiration date
Almond ELISA Kit II	23DESFAL002	NOV.30 2024

The product kit, batch 23DESFAL002, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : December 5, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.014	0.013	0.011	0.012	0.013	
0.5	0.058	0.056	0.056	0.056	0.057	1.8
5	0.466	0.459	0.465	0.457	0.462	1.0
10	0.895	0.891	0.911	0.901	0.900	1.0
25	2.076	2.060	2.040	2.052	2.057	0.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	7.9	7.3-13.5
B	7.3	7.2-13.3
C	7.5	5.9-10.9