



Morinaga Institute of Biological Science, Inc.

2-1-1 Shimosueyoshi, Tsurumi-ku,
Yokohama-shi, 230-8504, Japan
TEL:+81-45-586-2514
E-mail: info_miobs_e@morinaga.co.jp

April 14, 2023

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Almond ELISA Kit II	23APSFAL001	APR.13 2024

The product kit, batch 23APSFAL001, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : April 14, 2023)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.018	0.018	0.018	0.016	0.018	
0.5	0.060	0.056	0.057	0.055	0.057	3.8
5	0.487	0.462	0.464	0.454	0.467	3.0
10	0.954	0.896	0.893	0.867	0.903	4.1
25	2.139	2.092	2.050	2.071	2.088	1.8

(2) Results of the Test materials

Positive Control	Protein content($\mu\text{g/g}$)	Standard value
A	8.3	7.3-13.5
B	7.6	7.2-13.3
C	6.8	5.9-10.9