



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

March 15, 2022

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Hazelnut ELISA Kit II	22MASFHZ008	MAR.14 2023

The product kit, batch 22MASFHZ008, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : March 15, 2022)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.008	0.007	0.008	0.009	0.008	
0.39	0.057	0.057	0.055	0.053	0.056	3.5
0.78	0.099	0.097	0.094	0.094	0.096	2.6
1.56	0.174	0.176	0.171	0.172	0.173	1.3
3.13	0.339	0.322	0.325	0.320	0.327	2.6
6.25	0.639	0.628	0.615	0.613	0.624	1.9
12.5	1.230	1.210	1.180	1.159	1.195	2.6
25	2.209	2.183	2.164	2.121	2.169	1.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	5.4	3.2-6.0
B	4.6	2.7-5.1
C	4.8	3.0-5.6