



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

December 24, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Hazelnut ELISA Kit II	21DESFHZ007	DEC.23 2022

The product kit, batch 21DESFHZ007, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : December 24, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.012	0.011	0.009	0.010	0.011	
0.39	0.048	0.046	0.045	0.044	0.046	3.7
0.78	0.084	0.084	0.082	0.081	0.083	1.8
1.56	0.154	0.146	0.150	0.150	0.150	2.2
3.13	0.288	0.281	0.283	0.282	0.284	1.1
6.3	0.560	0.537	0.549	0.544	0.548	1.8
12.5	1.090	1.049	1.066	1.079	1.071	1.6
25	2.110	2.060	2.093	2.085	2.087	1.0

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	5.4	3.2-6.0
B	4.5	2.7-5.1
C	5.1	3.0-5.6