



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

September 16, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Hazelnut ELISA Kit II	21SESFHZ006	SEP.15 2022

The product kit, batch 21SESFHZ006, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : September 16, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.005	0.004	0.005	0.005	0.005	
0.39	0.040	0.039	0.040	0.040	0.040	1.3
0.78	0.071	0.073	0.071	0.077	0.073	3.9
1.56	0.137	0.136	0.135	0.138	0.137	0.9
3.13	0.260	0.259	0.263	0.269	0.263	1.7
6.3	0.488	0.488	0.488	0.503	0.492	1.5
12.5	0.945	0.937	0.941	0.967	0.948	1.4
25	1.727	1.729	1.762	1.796	1.754	1.9

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	4.8	3.2-6.0
B	4.2	2.7-5.1
C	4.5	3.0-5.6