



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

February 28, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Hazelnut ELISA Kit II	20FESFHZ003	FEB. 27. 2021

The product kit, batch 20FESFHZ003, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : February 28, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.006	0.007	0.007	0.006	0.007	
0.39	0.044	0.047	0.048	0.048	0.047	4.0
0.78	0.082	0.085	0.084	0.083	0.084	1.5
1.56	0.153	0.161	0.159	0.157	0.158	2.2
3.13	0.296	0.302	0.303	0.300	0.300	1.0
6.3	0.582	0.590	0.588	0.589	0.587	0.6
12.5	1.135	1.137	1.144	1.148	1.141	0.5
25	2.105	2.104	2.096	2.166	2.118	1.5

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	4.6	3.2-6.0
B	3.9	2.7-5.1
C	4.3	3.0-5.6