



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

September 20, 2019

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Hazelnut ELISA Kit II	19SESFHZ001A	SEP.05 2020

The product kit, batch 19SESFHZ001A, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : September 20, 2019)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.011	0.010	0.009	0.011	0.010	9.3
0.39	0.054	0.050	0.051	0.049	0.051	4.2
0.78	0.090	0.094	0.093	0.091	0.092	2.0
1.56	0.171	0.172	0.170	0.168	0.170	1.0
3.13	0.332	0.331	0.333	0.326	0.331	0.9
6.3	0.626	0.630	0.632	0.617	0.626	1.1
12.5	1.123	1.124	1.141	1.107	1.124	1.2
25	1.943	1.846	1.876	1.825	1.873	2.7

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	9.0	6.7-12.5
B	8.6	6.6-12.2
C	6.1	5.0-9.3