



CERTIFICATE OF ANALYSIS

LuciPac Pen

Product code :60331

Lot No.: 20240807U

Expiry date: Nov. 06, 2025

<u>TEST PARAMETER</u>	<u>RESULT</u>	<u>SPECIFICATION</u>
Average Luminescence(RLU)	12700	$\geq 9100, \leq 16900$

(When applied 100 $\mu$ L of 5 pmol AMP: measured with Lumitester Smart)

This is to certify that above information is a true and accurate copy of data obtained for this kit.

H Katayama  
Hiroshi Katayama; Manager

Date Aug. 23, 2024

KBG6033100ab-Eaa  
SA2408-0224

Quality Assurance Group, Edogawa  
Quality Unit  
Kikkoman Biochemifa Company