



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

September 3, 2021

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

M. Morita

Product:

Product name	Batch number	Expiration date
Wheat/Gluten (Gliadin) ELISA Kit II	21SESFGD094	SEP.02 2022

The product kit, batch 21SESFGD094, passed the quality control.

(1) Absorbance of the standard curve

(Assay date : September 3, 2021)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.004	0.004	0.004	0.004	0.004	
0.78	0.035	0.035	0.036	0.034	0.035	2.3
1.56	0.065	0.064	0.064	0.062	0.064	2.0
3.13	0.127	0.127	0.128	0.122	0.126	2.1
6.25	0.236	0.213	0.250	0.242	0.235	6.8
12.5	0.506	0.505	0.499	0.482	0.498	2.2
25	1.038	1.026	1.019	0.981	1.016	2.4
50	2.045	2.000	1.976	1.990	2.003	1.5

(2) Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.7	6.6-12.2
B	9.5	5.6-10.4
C	10.8	7.0-13.1