



Morinaga Institute of Biological Science, Inc.

SACHIURA 2-1-16, KANAZAWA-KU,
YOKOHAMA, 2360003, JAPAN
TEL:+81-45-791-7674
E-mail: info_miobs_e@morinaga.co.jp

November 10, 2020

Quality Control Report

Morinaga Institute of Biological Science, Inc.

Quality Assurance Department

Manager:

Product:

Product name	Batch number	Expiration date
Wheat/Gluten (Gliadin) ELISA Kit II	20NOSFGD081	NOV. 09 2021

The product kit, batch 20NOSFGD081, passed the quality control.

(1) Absorbance of the standard curve

(Assay date :November 10, 2020)

STD (ng/mL)	1	2	3	4	Ave.	CV (%)
0	0.010	0.008	0.009	0.005	0.008	
0.78	0.042	0.039	0.039	0.038	0.040	4.4
1.56	0.072	0.070	0.070	0.069	0.070	1.8
3.13	0.137	0.136	0.133	0.136	0.136	1.3
6.25	0.276	0.268	0.264	0.261	0.267	2.4
12.5	0.567	0.545	0.538	0.531	0.545	2.9
25	1.141	1.120	1.106	1.083	1.113	2.2
50	2.076	2.088	2.062	2.043	2.067	0.9

(2)Results of the Test materials

Positive Control	Protein content(μg/g)	Standard value
A	10.6	6.6-12.2
B	10.0	5.6-10.4
C	10.4	7.0-13.1