

Certificate of Analysis

Product	Compact Dry BC for detection of Bacillus cereus
ID-No.	1002970/ 1002971/ 1402971/ 1502991
Lot Number	028110
Expiry Date	31.03.2023

Test	Specifications	Results*
Appearance	Sheet of light yellow colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	6.8 –7.2	Corresponds
Colony growth		
<i>Bacillus cereus</i>	ATCC 19637 Blue-green colonies	Corresponds
<i>Bacillus cereus</i>	ATCC 11778 Blue-green colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633 No growth	Corresponds
<i>Enterococcus faecalis</i>	ATCC 19433 No growth	Corresponds
<i>Staphylococcus aureus</i>	ATCC 6538 No growth	Corresponds
<i>Escherichia coli</i>	ATCC 8739 No growth	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027 No growth	Corresponds

This lot meets the specifications.

Date: 02.08.2021

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.