

# Certificate of Analysis

**Product** Compact Dry TC for total counts  
**ID-No.** 1000166 / 1000167/ 1400168/ 1502879  
**Lot Number** 930111  
**Expiry Date** 31.10.2023

Test	Specifications	Results*
Appearance	Sheet of light yellow colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	6.8 –7.2	Corresponds
<b>Colony growth</b>		
<i>Escheria coli</i>	ATCC 8739 Red colonies	Corresponds
<i>Klebsiella pneumoniae</i>	ATCC 13883 Red colonies	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027 Red colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633 Red colonies	Corresponds
<i>Staphylococcus aureus</i>	ATCC 6538 Red colonies	Corresponds

**This lot meets the specifications.**

Date: 22.11.2021

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.