



Certificate of Analysis

Product Compact Dry TC for total counts
ID-No. 1000166 / 1000167/ 1400168/ 1502879
Lot Number 947205
Expiry Date 30.04.2024

Test	Specifications	Results*	
Appearance	Sheet of light yellow colour	Corresponds	
Loss on drying	< 10, 0 %	Corresponds	
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds	
pH value	6.8 –7.2	Corresponds	
Colony growth			
<i>Escheria coli</i>	ATCC 8739	Red colonies	Corresponds
<i>Klebsiella pneumoniae</i>	ATCC 13883	Red colonies	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027	Red colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633	Red colonies	Corresponds
<i>Staphylococcus aureus</i>	ATCC 6538	Red colonies	Corresponds

This lot meets the specifications.

Date: 13.08.2022

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.