

## Certificate of Analysis

**Product** Dilution rack for Compact Dry  
**ID-No.** 1 000 888  
**Lot Number** 0G5984  
**Expiry Date** 31.08.2022

Test	Specifications	Results*
<b>Aseptic test</b>	No growth of any bacteria	No growth
	Checked by PCA and PDA	Corresponds
<b>pH value</b>	7,1 ± 0,5	7,3
		Corresponds
<b>Appearance: Liquid</b>	Clear, slight yellow liquid	Clear, slight yellow liquid
		Corresponds
<b>Appearance: Bottle/Cup</b>	No visible particle	No visible particle
		Corresponds
<b>Labels</b>	Correct position; correct printing	Correct position; correct printing
<b>Net weight</b>	9,0 ± 0,3 g	9,1 g;
		Corresponds
<b>NaCl Content</b>	0,85 ± 0,1 %	0,9 %;
		Corresponds
<b>Leakage</b>	No leak; 0,03 Mpa for 30 minutes	No leakages
		Corresponds

**This lot meets the specifications.**

Date: 05.02.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.