

Certificate of Analysis

Product Dilution rack for Compact Dry
ID-No. 1 000 888
Lot Number 1G3384
Expiry Date 30.04.2023

Test	Specifications	Results*
Aseptic test	No growth of any bacteria	No growth
	Checked by PCA and PDA	Corresponds
pH value	7,1 ± 0,5	7,3
		Corresponds
Appearance: Liquid	Clear, slight yellow liquid	Clear, slight yellow liquid
		Corresponds
Appearance: Bottle/Cup	No visible particle	No visible particle
		Corresponds
Labels	Correct position; correct printing	Correct position; correct printing
Net weight	9,0 ± 0,3 g	9,1 g;
		Corresponds
NaCl Content	0,85 ± 0,1 %	0,9 %;
		Corresponds
Leakage	No leak; 0,03 Mpa for 30 minutes	No leakages
		Corresponds

This lot meets the specifications.

Date: 01.07.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.