

## Certificate of Analysis

**Product** Compact Dry BC  
**ID-No.** 1 002 970 / 1 002 971 / 1 002 972  
 1 402 991 / 1 402 991 / 1 502 991  
**Lot Number** 008102  
**Expiry Date** 31.07.2022

| Test                         | Specifications   | Results*    |
|------------------------------|--|-------------|
| Appearance                   | Sheet of light yellow colour                                     | Corresponds |
| Loss on drying               | < 10, 0 %  | Corresponds |
| Aseptic test                 | No growth of any bacteria.<br>(30°C for 5 days.<br>Visual check) | Corresponds |
| pH value                     | 6.8 -7.2   | Corresponds |
| <b>Colony growth</b>         |  |             |
| <i>Bacillus cereus</i>       | ATCC 19637 Blue-green colonies                                   | Corresponds |
| <i>Bacillus cereus</i>       | ATCC 11778 Blue-green colonies                                   | Corresponds |
| <i>Bacillus subtilis</i>     | ATCC 6633 No growth  | Corresponds |
| <i>Enterococcus faecalis</i> | ATCC 19433 No growth   | Corresponds |
| <i>Staphylococcus aureus</i> | ATCC 6538 No growth  | Corresponds |
| <i>Escherichia coli</i>      | ATCC 8739 No growth  | Corresponds |
| <i>Pseudomoas aeruginosa</i> | ATCC 9027 No growth  | Corresponds |

**This lot meets the specifications.**

Date: 04.03.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.