

Certificate of Analysis

Product Compact Dry BC
ID-No. 1 002 970 / 1 002 971 / 1 002 972
 1 402 991 / 1 402 991 / 1 502 991
Lot Number 015104
Expiry Date 30.09.2022

Test	Specifications	Results*
Appearance	Sheet of light yellow colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	6.8 -7.2	Corresponds
Colony growth		
<i>Bacillus cereus</i>	ATCC 19637 Blue-green colonies	Corresponds
<i>Bacillus cereus</i>	ATCC 11778 Blue-green colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633 No growth	Corresponds
<i>Enterococcus faecalis</i>	ATCC 19433 No growth	Corresponds
<i>Staphylococcus aureus</i>	ATCC 6538 No growth	Corresponds
<i>Escherichia coli</i>	ATCC 8739 No growth	Corresponds
<i>Pseudomoas aeruginosa</i>	ATCC 9027 No growth	Corresponds

This lot meets the specifications.

Date: 23.05.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.