

Certificate of Analysis

Product Compact Dry BC
ID-No. 1 002 970 / 1 002 971 / 1 002 972
1 402 991 / 1 402 991 / 1 502 991
Lot Number 026109
Expiry Date 28.02.2023

Test	Specifications	Results*	
Appearance	Sheet of light yellow colour	Corresponds	
Loss on drying	< 10, 0 %	Corresponds	
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds	
pH value	6.8 -7.2	Corresponds	
Colony growth			
<i>Bacillus cereus</i>	ATCC 19637	Blue-green colonies	Corresponds
<i>Bacillus cereus</i>	ATCC 11778	Blue-green colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633	No growth	Corresponds
<i>Enterococcus faecalis</i>	ATCC 19433	No growth	Corresponds
<i>Staphylococcus aureus</i>	ATCC 6538	No growth	Corresponds
<i>Escherichia coli</i>	ATCC 8739	No growth	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027	No growth	Corresponds

This lot meets the specifications.

Date: 03.12.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.