

Certificate of Analysis

HyServe
GmbH & Co. K.G.

Hechenrainer Str.24
82449 Uffing
am Staffelsee
Germany
☎ 0049-8846-1344
☎ 0049-8846-1342
www.HyServe.com
info@hyserve.com

Product Compact Dry ETB for Enterobacteriaceae
Ident-No. 1 002 941 /1 002 942 / 1 402 942/
1 502 889
Lot Number 261102
Expiry Date 31.05.2022

Test	Specifications	Results*
Appearance	Sheet of light pink colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	7.2 - 7.6	Corresponds
Colony growth		
<i>Citrobacter freundii</i>	Reddish purple colonies	Corresponds
ATCC8090	Reddish purple colonies	Corresponds
<i>Klebsiella oxytoca</i> ATCC13182	Reddish purple colonies	Corresponds
<i>Escherichia coli</i> ATCC8739	Reddish purple colonies	Corresponds
<i>Salmonella typhimurium</i> ATCC 13311	Reddish purple colonies	Corresponds
<i>Proteus mirabilis</i> ATCC29906		
<i>Bacillus subtilis</i> ATCC6633	No growth	
<i>Staphylococcus aureus</i> ATCC6538	No growth	

This lot meets the specifications.

Date: 13.03.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.