HyServe
GmbH & Co. K.G.

Hechenrainer Str.24
82449 Uffing
am Staffelsee
Germany
☎ 0049-8846-1344
🗐 0049-8846-1342
www.HyServe.com
info@hyserve.com

Certificate of Analysis

Product	**Compact Dry ETB for Enterobacteriacae**
Ident-No.	**1 002 941 /1 002 942 / 1 402 942/ 1 502 889**
Lot Number	**266104**
Expiry Date	**31.07.2022**

Test		**Specifications**	**Results***
Appearance		Sheet of light pink colour	Corresponds
Loss on drying		< 10, 0 %	Corresponds
Aseptic test		No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value		7.2 – 7.6	Corresponds
Colony growth			
Citrobacter freundii		Reddish purple colonies	Corresponds
ATCC8090		Reddish purple colonies	Corresponds
Klebsiella oxytoca	ATCC13182	Reddish purple colonies	Corresponds
Escherichia coli	ATCC8739	Reddish purple colonies	Corresponds
Salmonella typhimurium		Reddish purple colonies	Corresponds
	ATCC 13311		
Proteus mirabilis	ATCC29906		
Bacillus subtilis	ATCC6633	No growth	
Staphylococcus aureus	ATCC6538	No growth	

This lot meets the specifications.

Date: 19.05.2021 Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.