

# Certificate of Analysis

**HyServe**  
GmbH & Co. K.G.

Hechenrainer Str.24  
82449 Uffing  
am Staffelsee  
Germany  
☎ 0049-8846-1344  
☎ 0049-8846-1342  
[www.HyServe.com](http://www.HyServe.com)  
[info@hyserve.com](mailto:info@hyserve.com)

**Product** Compact Dry ETB for Enterobacteriaceae  
**Ident-No.** 1 002 941 /1 002 942 / 1 402 942/  
1 502 889  
**Lot Number** 269106  
**Expiry Date** 30.09.2022

Test	Specifications	Results*
Appearance	Sheet of light pink colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. ( 30°C for 5 days. Visual check )	Corresponds
pH value	7.2 – 7.6	Corresponds
<b>Colony growth</b>		
<i>Citrobacter freundii</i>	Reddish purple colonies	Corresponds
ATCC8090	Reddish purple colonies	Corresponds
<i>Klebsiella oxytoca</i> ATCC13182	Reddish purple colonies	Corresponds
<i>Escherichia coli</i> ATCC8739	Reddish purple colonies	Corresponds
<i>Salmonella typhimurium</i> ATCC 13311	Reddish purple colonies	Corresponds
<i>Proteus mirabilis</i> ATCC29906		
<i>Bacillus subtilis</i> ATCC6633	No growth	
<i>Staphylococcus aureus</i> ATCC6538	No growth	

**This lot meets the specifications.**

Date: 11.06.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.