

Certificate of Analysis

Product Compact Dry PA for Pseudomonas aeruginosa
ID-No. 1 002 945 / 1 002 947 / 1002 948 / 1 502 949
Lot Number 031111
Expiry Date 31.05.2022

Test	Specifications	Judge*
Appearance	White to very light pink	Pass
Loss on drying	<20%	Pass
Aseptic test	No growth of any bacteria.	Pass
pH value	7.2 +/- 0.2	Pass

Evaluation of Performance

Specification

- When following test strains are inoculated to incubate at 35°C for 24 hours good grow and color development should be observed:

Test strain	Growth	Color development
Pseudomonas aeruginosa NBRC 13275	+	red
Pseudomonas aeruginosa ATCC 27853	+	red
Pseudomonas aeruginosa ATCC 10145	+	red

- When following test strains are inoculated to incubate at 35°C for 24 hours growths should be inhibited:

Test strain
 Escherichia coli NBRC 3972
 Bacillus subtilis ATCC 6633
 Staphylococcus aureus ATCC 6538

Test results

Good growths were observed for the above strains (1).
 Inhibitions were observed for the below strains (2).

This lot meets the specification.

Date: 07.12.2021

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.