

Certificate of Analysis

Product Compact Dry TC for total counts
ID-No. 1 000 166 / 1 000 167 / 1 400 168/
1 502 879
Lot Number 744106
Expiry Date 31.05.2023

Test	Specifications	Results*
Appearance	Sheet of light yellow colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	6.8 -7.2	Corresponds
Colony growth		
<i>Escheria coli</i>	ATCC 8739 Red colonies	Corresponds
<i>Klebsiella pneumoniae</i>	ATCC 13883 Red colonies	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027 Red colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633 Red colonies	Corresponds
<i>Staphylococcus aureus</i>	ATCC 6538 Red colonies	Corresponds

This lot meets the specifications.

Date: 15.10.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.