

## Certificate of Analysis

**Product** Compact Dry TC for total counts  
**ID-No.** 1 000 166 / 1 000 167 / 1 400 168/  
1 502 879  
**Lot Number** 745107  
**Expiry Date** 30.06.2023

Test	Specifications	Results*
<b>Appearance</b>	Sheet of light yellow colour	Corresponds
<b>Loss on drying</b>	< 10, 0 %	Corresponds
<b>Aseptic test</b>	No growth of any bacteria. ( 30°C for 5 days. Visual check )	Corresponds
<b>pH value</b>	6.8 -7.2	Corresponds
<b>Colony growth</b>		
<i>Escheria coli</i>	ATCC 8739 Red colonies	Corresponds
<i>Klebsiella pneumoniae</i>	ATCC 13883 Red colonies	Corresponds
<i>Pseudomonas aeruginosa</i>	ATCC 9027 Red colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633 Red colonies	Corresponds
<i>Staphylococcus aureus</i>	ATCC 6538 Red colonies	Corresponds

**This lot meets the specifications.**

Date: 28.07.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.