

# Certificate of Analysis

**HyServe**  
GmbH & Co. K.G.

Hechenrainerstr.12  
82449 Uffing  
Germany

**Product** Compact Dry YMR (rapid)  
for yeast and moulds  
**ID-No.** 1 600 869/ 1 600 870 / 1 600 871  
1 602 882  
**Lot Number** 074107  
**Expiry Date** 30.06.2023

📞 0049-8846-1344  
📠 0049-8846-1342  
[www.HyServe.com](http://www.HyServe.com)  
[info@hyserve.com](mailto:info@hyserve.com)

Test	Specifications	Results*
<b>Appearance</b>	Sheet of light yellow colour	Corresponds
<b>Loss on drying</b>	< 10, 0 %	Corresponds
<b>Aseptic test</b>	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
<b>pH value</b>	5.2 – 5.8	Corresponds
<b>Colony growth</b>		
<i>Candida glabrata</i>	JCM 3699 Light bluish green colonies	Corresponds
<i>Candida albicans</i>	ATCC10231 Light yellow green colonies	Corresponds
<i>Aspergillus brasiliensis</i>	ATCC 16404 Blue colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633 No growth	Corresponds
<i>Escherichia coli</i>	ATCC 8739 No growth	Corresponds

**This lot meets the specifications.**

Date: 20.09.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.