

## Certificate of Analysis

**Product** Compact Dry YM for yeast and moulds  
**ID-No.** 1 000 869/ 1 000 870 / 1 400 871 /  
1 502 882  
**Lot Number** 515108  
**Expiry Date** 31.01.2023

Test	Specifications	Results*	
Appearance	Sheet of light yellow colour	Corresponds	
Loss on drying	< 10, 0 %	Corresponds	
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds	
pH value	5.2 – 5.8	Corresponds	
<b>Colony growth</b>			
<i>Candida albicans</i> ATCC 2091	Light bluish green colonies	Corresponds	
<i>Candida albicans</i> ATCC 10231	Light bluish green colonies	Corresponds	
<i>Aspergillus niger</i> ATCC 16404	Blue colonies	Corresponds	
<i>Bacillus subtilis</i> ATCC 6633	No growth	Corresponds	
<i>Escherichia coli</i> ATCC 8739	No growth	Corresponds	

**This lot meets the specifications.**

Date: 19.12.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.