

Certificate of Analysis

Product Compact Dry YM for yeast and moulds
ID-No. 1 000 869/ 1 000 870 / 1 400 871 /
1 502 882
Lot Number 518109
Expiry Date 28.02.2023

Test	Specifications	Results*
Appearance	Sheet of light yellow colour	Corresponds
Loss on drying	< 10, 0 %	Corresponds
Aseptic test	No growth of any bacteria. (30°C for 5 days. Visual check)	Corresponds
pH value	5.2 – 5.8	Corresponds
Colony growth		
<i>Candida albicans</i> ATCC 2091	Light bluish green colonies	Corresponds
<i>Candida albicans</i> ATCC 10231	Light bluish green colonies	Corresponds
<i>Aspergillus niger</i> ATCC 16404	Blue colonies	Corresponds
<i>Bacillus subtilis</i> ATCC 6633	No growth	Corresponds
<i>Escherichia coli</i> ATCC 8739	No growth	Corresponds

This lot meets the specifications.

Date: 19.12.2021

Dr. Stephan Speidel

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.