

## Certificate of Analysis

**Product** Compact Dry YM for yeast and moulds  
**ID-No.** 1000869/ 1000870/ 1400871/ 1502882  
**Lot Number** 520110  
**Expiry Date** 31.03.2023

Test	Specifications	Results*	
<b>Appearance</b>	Sheet of light yellow colour	Corresponds	
<b>Loss on drying</b>	< 10, 0 %	Corresponds	
<b>Aseptic test</b>	No growth of any bacteria.  (30°C for 5 days. Visual check)	Corresponds	
<b>pH value</b>	5.2 – 5.8	Corresponds	
<b>Colony growth</b>			
<i>Candida albicans</i>	ATCC 2091	Light bluish green colonies	Corresponds
<i>Candida albicans</i>	ATCC 10231	Light bluish green colonies	Corresponds
<i>Aspergillus niger</i>	ATCC 16404	Blue colonies	Corresponds
<i>Bacillus subtilis</i>	ATCC 6633	No growth	Corresponds
<i>Escherichia coli</i>	ATCC 8739	No growth	Corresponds

**This lot meets the specifications.**

Date: 29.04.2022

This product is produced in the ISO 9001 certified production facility.

\*Data is obtained either by OEM partner or HyServe.