

Certificate of Analysis

Product Compact Dry AQ
ID-No. 1 002 950/1 002 951/ 1 402 951/
 1 502 992
Lot Number 121107
Expiry Date 30.06.2023

Test	Specifications	Results*
Appearance	Light pink, no particular matters	Corresponds
Loss on drying	Less than 10.0 %	Corresponds
Aseptic test	No growth of any bacteria.	Corresponds
pH value	7.0 – 7.4	Corresponds

Evaluation of Performance, Specification

Test strain	Growth	Color
Bacillus subtilis ATCC6633	+	red
Staphylococcus aureus ATCC6538	+	red
Escherichia coli ATCC8793	+	red
Pseudomonas aeruginosa ATCC9027	+	red
Pseudomonas fluorescens NBRC 15842	+	white-pink
Candida albicans ATCC10231		

Test results

Good growths were observed for the above strains.

Date: 05.11.2021

This product is produced in the ISO 9001 certified production facility.

*Data is obtained either by OEM partner or HyServe.